



LEGISLATIVE ASSEMBLY FOR THE AUSTRALIAN CAPITAL TERRITORY

SELECT COMMITTEE ON ESTIMATES 2014-2015
Brendan Smyth MLA (Chair), Mary Porter MLA (Deputy-Chair),
Giulia Jones MLA, Yvette Berry MLA

ANSWER TO QUESTION TAKEN ON NOTICE DURING PUBLIC HEARINGS

Asked by Mr Brendan Smyth MLA on 26 June 2014: Mr Simon Corbell MLA took on notice the following question(s):

[Ref: Hansard Transcript [26 June 2014] [PAGE 87]]

In relation to: MRX monitors (JACS #16)

THE CHAIR: All right. And was there an international recall of the MRX monitors or batteries?

Mr Lane: Not to my knowledge, Mr Chair, but I would have to take that question on notice.

Mr Simon Corbell MLA: The answer to the Member's question is as follows:-

Worldwide recalls issued by Phillips (the manufacturer) have been for the MRx monitor batteries and MRx monitors with serial numbers within the range of US00100204 to US005790921.

43,792 MRx monitor batteries were recalled of which the ACT had 137.

3 of the 56 MRx monitors purchased by the ACT fell within the serial number range stated above, and were fixed by the supplier.

Approved for circulation to the Select Committee on Estimates 2014-2015

Signature:

Date:

By the Minister for Police and Emergency Services, Mr Simon Corbell MLA